



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

PSG CENTRE FOR INDUSTRIAL RESEARCH AND DEVELOPMENT, A 115, PSG
COLLEGE OF TECHNOLOGY, AVINASHI ROAD, PEELAMEDU, COIMBATORE, TAMIL
NADU, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-4660

Page No

1 of 26

Validity

30/01/2026 to 29/01/2030

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
Permanent Facility					
1	MEDICAL DEVICES- DISCHARGE EQUIPMENT/DE VICES	BP Apparatus - Pulse Rate	Using Vital Sign Simulator by Direct Method	30 bpm to 150 bpm	4.22 bpm
2	MEDICAL DEVICES- DISCHARGE EQUIPMENT/DE VICES	BP Apparatus / Sphygmomanometer - Pressure (Leak Test)	Using Vital Sign Simulator by Direct Method	0 to 40 mmHg	5.8 mmHg
3	MEDICAL DEVICES- DISCHARGE EQUIPMENT/DE VICES	BP Apparatus / Sphygmomanometer - Pressure (NIBP)	Using Vital Sign Simulator by Direct Method	50 mmHg to 200 mmHg	5.8 mmHg
4	MEDICAL DEVICES- DISCHARGE EQUIPMENT/DE VICES	Flow Meter with Humidifier - Pressure	Using Gas Flow Analyzer by Comparison Method	0 to 60 psi	1.95 %
5	MEDICAL DEVICES- DISCHARGE EQUIPMENT/DE VICES	Flow Meter with Humidifier - Temperature	Using Digital Thermometer by Comparison Method	20 °C to 50 °C	1 °C



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

PSG CENTRE FOR INDUSTRIAL RESEARCH AND DEVELOPMENT, A 115, PSG
COLLEGE OF TECHNOLOGY, AVINASHI ROAD, PEELAMEDU, COIMBATORE, TAMIL
NADU, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-4660

Page No

2 of 26

Validity

30/01/2026 to 29/01/2030

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
6	MEDICAL DEVICES-DISCHARGE EQUIPMENT/DEVICES	Flow Meter with Humidifier / Flow Meter / Oxygen Concentrator - Flow	Using Gas Flow Analyzer by Comparison Method	1 lpm to 15 lpm	5.2 %
7	MEDICAL DEVICES-DISCHARGE EQUIPMENT/DEVICES	Infusion Pump / Feeding Pump - Flow	Using Infusion Device Analyzer by Direct Method	5 ml/hr to 1000 ml/hr	4.28 %
8	MEDICAL DEVICES-DISCHARGE EQUIPMENT/DEVICES	Infusion Pump / Feeding Pump - Occlusion Pressure	Using Infusion Device Analyzer by Direct Method	0 to 1300 mmHg	2.07 %
9	MEDICAL DEVICES-DISCHARGE EQUIPMENT/DEVICES	Infusion Pump / Feeding Pump - Volume	Using Infusion Device Analyzer by Direct Method	1 ml to 400 ml	6.4 %
10	MEDICAL DEVICES-DISCHARGE EQUIPMENT/DEVICES	Nebulizer - Flow	Using Gas Flow Analyzer by Comparison Method	7 lpm to 10 lpm	5 %
11	MEDICAL DEVICES-DISCHARGE EQUIPMENT/DEVICES	Nebulizer - Pressure	Using Gas Flow Analyzer by Comparison Method	0 to 50 psi	4.97 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

PSG CENTRE FOR INDUSTRIAL RESEARCH AND DEVELOPMENT, A 115, PSG
COLLEGE OF TECHNOLOGY, AVINASHI ROAD, PEELAMEDU, COIMBATORE, TAMIL
NADU, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-4660

Page No

3 of 26

Validity

30/01/2026 to 29/01/2030

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
12	MEDICAL DEVICES-DISCHARGE EQUIPMENT/DEVICES	Pulse Oximeter - Heart Rate	Using SPO2 Functional Tester by Direct Method	30 bpm to 150 bpm	13.07 %
13	MEDICAL DEVICES-DISCHARGE EQUIPMENT/DEVICES	Pulse Oximeter - SPO2	Using SPO2 Functional Tester by Direct Method	80 % to 100 %	4.46 %
14	MEDICAL DEVICES-DISCHARGE EQUIPMENT/DEVICES	Suction Pump - Flow	Using Gas Flow Analyzer by Comparison Method	1 lpm to 15 lpm	2.41 %
15	MEDICAL DEVICES-DISCHARGE EQUIPMENT/DEVICES	Suction Pump - Vacuum	Using Gas Flow Analyzer by Comparison Method	0 to 600 mmHg	13.8 mmHg
16	MEDICAL DEVICES-DISCHARGE EQUIPMENT/DEVICES	Syringe Pump - Flow	Using Infusion Device Analyzer by Direct method	5 ml/hr to 1000 ml/hr	4.28 %
17	MEDICAL DEVICES-DISCHARGE EQUIPMENT/DEVICES	Syringe Pump - Occlusion Pressure	Using Infusion Device Analyzer by Direct Method	0 to 900 mmHg	4.29 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

PSG CENTRE FOR INDUSTRIAL RESEARCH AND DEVELOPMENT, A 115, PSG
COLLEGE OF TECHNOLOGY, AVINASHI ROAD, PEELAMEDU, COIMBATORE, TAMIL
NADU, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-4660

Page No

4 of 26

Validity

30/01/2026 to 29/01/2030

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
18	MEDICAL DEVICES-DISCHARGE EQUIPMENT/DEVICES	Syringe Pump - Volume	Using Infusion Device Analyzer by Direct Method	1 ml to 50 ml	6.4 %
19	MEDICAL DEVICES-DISCHARGE EQUIPMENT/DEVICES	Syringe Pump, Infusion Pump, Nebulizer, Pulse Oximeter, BP Apparatus, Suction Pump - Chassis/Enclosure Leakage Current	Using Electrical Safety Analyzer by Direct Method	10 μ A to 9000 μ A	3.6 μ A to 223 μ A
20	MEDICAL DEVICES-DISCHARGE EQUIPMENT/DEVICES	Syringe Pump, Infusion Pump, Nebulizer, Pulse Oximeter, BP Apparatus, Suction Pump - Earth Leakage Current	Using Electrical Safety Analyzer by Direct Method	10 μ A to 9000 μ A	3.6 μ A to 223 μ A
21	MEDICAL DEVICES-DISCHARGE EQUIPMENT/DEVICES	Syringe Pump, Infusion Pump, Nebulizer, Pulse Oximeter, BP Apparatus, Suction Pump - Insulation Resistance Test @ 500 V	Using Electrical Safety Analyzer by Direct Method	1 Mohm to 100 Mohm	8.98 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

PSG CENTRE FOR INDUSTRIAL RESEARCH AND DEVELOPMENT, A 115, PSG
COLLEGE OF TECHNOLOGY, AVINASHI ROAD, PEELAMEDU, COIMBATORE, TAMIL
NADU, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-4660

Page No

5 of 26

Validity

30/01/2026 to 29/01/2030

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
22	MEDICAL DEVICES-DISCHARGE EQUIPMENT/DEVICES	Syringe Pump, Infusion Pump, Nebulizer, Pulse Oximeter, BP Apparatus, Suction Pump - Patient Leakage Current	Using Electrical Safety Analyzer by Direct Method	10 μ A to 1000 μ A	3.6 μ A to 24.5 μ A
23	MEDICAL DEVICES-DISCHARGE EQUIPMENT/DEVICES	Syringe Pump, Infusion Pump, Nebulizer, Pulse Oximeter, BP Apparatus, Suction Pump - Protective Earth Resistance	Using Electrical Safety Analyzer by Direct Method	0.1 ohm to 1.9 ohm	0.022 ohm to 0.058 ohm
24	MEDICAL DEVICES-IMAGING/PLOTTERS	ECG Machine - Chassis / Enclosure Leakage Current	Using Electrical Safety Analyzer by Direct Method	10 μ A to 9000 μ A	3.6 μ A to 223 μ A
25	MEDICAL DEVICES-IMAGING/PLOTTERS	ECG Machine - Earth Leakage Current	Using Electrical Safety Analyzer by Direct Method	10 μ A to 10 mA	3.6 μ A to 223 μ A
26	MEDICAL DEVICES-IMAGING/PLOTTERS	ECG Machine - Insulation Resistance Test @ 500 V	Using Electrical Safety Analyzer by Direct Method	1 Mohm to 100 Mohm	3 %
27	MEDICAL DEVICES-IMAGING/PLOTTERS	ECG Machine - Patient Leakage Current	Using Electrical Safety Analyzer by Direct Method	10 μ A to 1000 μ A	3.6 μ A to 24.5 μ A



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

PSG CENTRE FOR INDUSTRIAL RESEARCH AND DEVELOPMENT, A 115, PSG
COLLEGE OF TECHNOLOGY, AVINASHI ROAD, PEELAMEDU, COIMBATORE, TAMIL
NADU, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-4660

Page No

6 of 26

Validity

30/01/2026 to 29/01/2030

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
28	MEDICAL DEVICES-IMAGING/PLOTTERS	ECG Machine - Protective Earth Resistance	Using Electrical Safety Analyzer by Direct Method	0.05 ohm to 1.9 ohm	0.022 ohm to 0.058 ohm
29	MEDICAL DEVICES-IMAGING/PLOTTERS	ECG Machine / Recorder - ECG Output Amplitude	Using Defibrillator Analyzer by Direct Method	0.5 mV to 5 mV	4.8 %
30	MEDICAL DEVICES-IMAGING/PLOTTERS	Heart Rate - ECG Machine/Recorder, Monitors	Using Defibrillator Analyzer by Direct Method	30 bpm to 300 bpm	2.27 %
31	MEDICAL DEVICES-MONITORING UNIT	Patient Monitor / Multipara Monitor - Chassis / Enclosure Leakage Current	Using Electrical Safety Analyzer by Direct Method	10 μ A to 9000 μ A	3.6 μ A to 223 μ A
32	MEDICAL DEVICES-MONITORING UNIT	Patient Monitor / Multipara Monitor - Earth Leakage Current	Using Electrical Safety Analyser by Direct Method	10 μ A to 9000 μ A	3.6 μ A to 223 μ A
33	MEDICAL DEVICES-MONITORING UNIT	Patient Monitor / Multipara Monitor - ECG	Using Vital Sign Simulator by Direct Method	0.05 mV to 5 mV	6.26 %
34	MEDICAL DEVICES-MONITORING UNIT	Patient Monitor / Multipara Monitor - Heart Rate	Using SPO2 Simulator and Vital Sign Simulator by Direct Method	30 BPM to 300 BPM	2.27 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

PSG CENTRE FOR INDUSTRIAL RESEARCH AND DEVELOPMENT, A 115, PSG
COLLEGE OF TECHNOLOGY, AVINASHI ROAD, PEELAMEDU, COIMBATORE, TAMIL
NADU, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-4660

Page No

7 of 26

Validity

30/01/2026 to 29/01/2030

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
35	MEDICAL DEVICES-MONITORING UNIT	Patient Monitor / Multipara Monitor - Invasive Blood Pressure Accuracy	Using SPO2 Simulator and Vital Sign Simulator by Direct Method	0 to 200 mmHg	5.33 %
36	MEDICAL DEVICES-MONITORING UNIT	Patient Monitor / Multipara Monitor - Non Invasive Blood Pressure	Using Vital Sign Simulator by Direct Method	50 to 400 mmHg	5.8 mmHg
37	MEDICAL DEVICES-MONITORING UNIT	Patient Monitor / Multipara Monitor - Non Invasive Blood Pressure (Leak Test)	Using Vital Sign Simulator by Direct Method	50 mmHg to 400 mmHg	1.3 mmHg
38	MEDICAL DEVICES-MONITORING UNIT	Patient Monitor / Multipara Monitor - Patient Leakage Current	Using Electrical Safety Analyzer by Direct Method	10 μ A to 1000 μ A	3.6 μ A to 24.5 μ A
39	MEDICAL DEVICES-MONITORING UNIT	Patient Monitor / Multipara Monitor - Protective Earth Resistance	Using Electrical Safety Analyzer by Direct Method	0.05 ohm to 1.9 ohm	0.022 ohm to 0.058 ohm
40	MEDICAL DEVICES-MONITORING UNIT	Patient Monitor / Multipara Monitor - Respiration Rate	Using Vital Sign Simulator by Direct Method	10 brpm to 100 brpm	6.6 %
41	MEDICAL DEVICES-MONITORING UNIT	Patient Monitor / Multipara Monitor - SPO2	Using SPO2 Simulator by Direct Method	70 % to 100 %	4.48 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

PSG CENTRE FOR INDUSTRIAL RESEARCH AND DEVELOPMENT, A 115, PSG
COLLEGE OF TECHNOLOGY, AVINASHI ROAD, PEELAMEDU, COIMBATORE, TAMIL
NADU, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-4660

Page No

8 of 26

Validity

30/01/2026 to 29/01/2030

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
42	MEDICAL DEVICES- MONITORING UNIT	Patient Monitor, Syringe / Infusion Pump, Defibrillator, Suction Pump, ECG Unit, Electro - Surgical Unit, Autoclave, Radiant Warmer, Patient Warmer, Infant Incubator - Insulation Resistance @ 500 V	Using Electrical Safety Analyzer by Direct Method	1 Mohm to 100 Mohm	8.98 %
43	MEDICAL DEVICES- PATIENT CONDITIONING / MAINTENANCE	Autoclave - Pressure	Using Digital Pressure Gauge by Comparison Method	0 to 2 bar	0.1 bar
44	MEDICAL DEVICES- PATIENT CONDITIONING / MAINTENANCE	Autoclave - Temperature	Using Digital Thermometer with Logger by Comparison Method	50 °C to 125 °C	1.3 °C
45	MEDICAL DEVICES- PATIENT CONDITIONING / MAINTENANCE	Autoclave - Time	Using Digital Timer by Comparison Method	1 s to 45 Min	0.1 s to 10 s



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

PSG CENTRE FOR INDUSTRIAL RESEARCH AND DEVELOPMENT, A 115, PSG
COLLEGE OF TECHNOLOGY, AVINASHI ROAD, PEELAMEDU, COIMBATORE, TAMIL
NADU, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-4660

Page No

9 of 26

Validity

30/01/2026 to 29/01/2030

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
46	MEDICAL DEVICES- PATIENT CONDITIONING / MAINTENANCE	Defibrillator - Charge Time	Using Defibrillator Analyzer by Direct Method	1.5 s to 10 s	0.17 s
47	MEDICAL DEVICES- PATIENT CONDITIONING / MAINTENANCE	Defibrillator - ECG Amplitude	Using Defibrillator Analyzer by Direct Method	0.5 mV to 5 mV	6.26 %
48	MEDICAL DEVICES- PATIENT CONDITIONING / MAINTENANCE	Defibrillator - Energy	Using Defibrillator Analyzer by Direct Method	20 J to 300 J	7.5 J
49	MEDICAL DEVICES- PATIENT CONDITIONING / MAINTENANCE	Defibrillator - Heart Rate	Using Defibrillator Analyzer by Direct Method	30 bpm to 300 bpm	1.6 % to 2.27 %
50	MEDICAL DEVICES- PATIENT CONDITIONING / MAINTENANCE	Defibrillator - NIBP	Using Vital Signs Simulator by Direct Method	50 mmHg to 200 mmHg	5.8 mmHg



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

PSG CENTRE FOR INDUSTRIAL RESEARCH AND DEVELOPMENT, A 115, PSG
COLLEGE OF TECHNOLOGY, AVINASHI ROAD, PEELAMEDU, COIMBATORE, TAMIL
NADU, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-4660

Page No

10 of 26

Validity

30/01/2026 to 29/01/2030

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
51	MEDICAL DEVICES- PATIENT CONDITIONING / MAINTENANCE	Defibrillator - SPO2	Using SPO2 Functional Tester by Direct Method	80 % to 100 %	4.46 %
52	MEDICAL DEVICES- PATIENT CONDITIONING / MAINTENANCE	Defibrillator, Infant Incubator, Electro Surgical Unit / Diathermy Machine / Cautery Machine, Autoclave, Patient Warmer, Radiant Warmer - Chassis / Enclosure Leakage Current	Using Electrical Safety Analyzer by Direct Method	10 μ A to 9000 μ A	3.6 μ A to 223 μ A
53	MEDICAL DEVICES- PATIENT CONDITIONING / MAINTENANCE	Defibrillator, Infant Incubator, Electro Surgical Unit / Diathermy Machine / Cautery Machine, Autoclave, Patient Warmer, Radiant Warmer - Earth Leakage Current	Using Electrical Safety Analyzer by Direct Method	10 μ A to 9000 μ A	3.6 μ A to 223 μ A



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

PSG CENTRE FOR INDUSTRIAL RESEARCH AND DEVELOPMENT, A 115, PSG
COLLEGE OF TECHNOLOGY, AVINASHI ROAD, PEELAMEDU, COIMBATORE, TAMIL
NADU, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-4660

Page No

11 of 26

Validity

30/01/2026 to 29/01/2030

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
54	MEDICAL DEVICES- PATIENT CONDITIONING / MAINTENANCE	Defibrillator, Infant Incubator, Electro Surgical Unit / Diathermy Machine / Cautery Machine, Autoclave, Patient Warmer, Radiant Warmer - Insulation Resistance Test @ 500 V	Using Electrical Safety Analyzer by Direct Method	1 Mohm to 100 Mohm	8.98 %
55	MEDICAL DEVICES- PATIENT CONDITIONING / MAINTENANCE	Defibrillator, Infant Incubator, Electro Surgical Unit / Diathermy Machine / Cautery Machine, Autoclave, Patient Warmer, Radiant Warmer - Patient Leakage Current	Using Electrical Safety Analyzer by Direct Method	10 μ A to 1000 μ A	3.6 μ A to 24.5 μ A
56	MEDICAL DEVICES- PATIENT CONDITIONING / MAINTENANCE	Defibrillator, Infant Incubator, Electro Surgical Unit / Diathermy Machine / Cautery Machine, Autoclave, Patient Warmer, Radiant Warmer - Protective Earth Resistance	Using Electrical Safety Analyser by Direct Method	0.1 ohm to 1.9 ohm	0.022 ohm to 0.058 ohm



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

PSG CENTRE FOR INDUSTRIAL RESEARCH AND DEVELOPMENT, A 115, PSG
COLLEGE OF TECHNOLOGY, AVINASHI ROAD, PEELAMEDU, COIMBATORE, TAMIL
NADU, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-4660

Page No

12 of 26

Validity

30/01/2026 to 29/01/2030

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
57	MEDICAL DEVICES- PATIENT CONDITIONING / MAINTENANCE	Electro Surgical Unit / Diathermy Machine / Cautery Machine - HF Leakage	Using Electro Surgery Analyzer by Direct Method	11 mA to 150 mA	1.62 mA to 5.64 mA
58	MEDICAL DEVICES- PATIENT CONDITIONING / MAINTENANCE	Electro Surgical Unit / Diathermy Machine / Cautery Machine - Power	Using Electro Surgery Analyzer by Direct Method	10 W to 300 W	8.48 %
59	MEDICAL DEVICES- PATIENT CONDITIONING / MAINTENANCE	Electro Surgical Unit / Diathermy Machine / Cautery Machine - REM	Using Electro Surgery Analyzer by Direct Method	10 ohm to 475 ohm	3.08 %
60	MEDICAL DEVICES- PATIENT CONDITIONING / MAINTENANCE	Pacemaker / Defibrillator - Pacer Output	Using Defibrillator analyzer by Direct Method	2 mA to 20 mA	3.8 %
61	MEDICAL DEVICES- PATIENT CONDITIONING / MAINTENANCE	Pacemaker / Defibrillator - Pacer Rate	Using Defibrillator analyzer by Direct Method	30 ppm to 180 ppm	1.92 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

PSG CENTRE FOR INDUSTRIAL RESEARCH AND DEVELOPMENT, A 115, PSG
COLLEGE OF TECHNOLOGY, AVINASHI ROAD, PEELAMEDU, COIMBATORE, TAMIL
NADU, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-4660

Page No

13 of 26

Validity

30/01/2026 to 29/01/2030

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
62	MEDICAL DEVICES- PATIENT CONDITIONING / MAINTENANCE	Patient Warmer - Temperature	Using Digital Thermometer with Logger by Comparison Method	30 °C to 45 °C	1 °C
63	MEDICAL DEVICES- PATIENT CONDITIONING / MAINTENANCE	Radiant Warmer - Temperature	Using Digital Thermometer with Logger by Comparison Method	30 °C to 40 °C	1.2 °C



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

PSG CENTRE FOR INDUSTRIAL RESEARCH AND DEVELOPMENT, A 115, PSG
COLLEGE OF TECHNOLOGY, AVINASHI ROAD, PEELAMEDU, COIMBATORE, TAMIL
NADU, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-4660

Page No

14 of 26

Validity

30/01/2026 to 29/01/2030

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)($\pm$)
Site Facility					
1	MEDICAL DEVICES- DISCHARGE EQUIPMENT/DE VICES	BP Apparatus - Pulse Rate	Using Vital Sign Simulator by Direct Method	30 bpm to 150 bpm	4.22 bpm
2	MEDICAL DEVICES- DISCHARGE EQUIPMENT/DE VICES	BP Apparatus / Sphygmomanometer - Pressure (Leak Test)	Using Vital Sign Simulator by Direct Method	0 to 40 mmHg	5.8 mmHg
3	MEDICAL DEVICES- DISCHARGE EQUIPMENT/DE VICES	BP Apparatus / Sphygmomanometer - Pressure (NIBP)	Using Vital Sign Simulator by Direct Method	50 mmHg to 200 mmHg	5.8 mmHg
4	MEDICAL DEVICES- DISCHARGE EQUIPMENT/DE VICES	Flow Meter with Humidifier - Pressure	Using Gas Flow Analyzer by Comparison Method	0 to 60 psi	1.95 %
5	MEDICAL DEVICES- DISCHARGE EQUIPMENT/DE VICES	Flow Meter with Humidifier - Temperature	Using Digital Thermometer by Comparison Method	20 °C to 50 °C	1 °C



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

PSG CENTRE FOR INDUSTRIAL RESEARCH AND DEVELOPMENT, A 115, PSG
COLLEGE OF TECHNOLOGY, AVINASHI ROAD, PEELAMEDU, COIMBATORE, TAMIL
NADU, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-4660

Page No

15 of 26

Validity

30/01/2026 to 29/01/2030

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
6	MEDICAL DEVICES-DISCHARGE EQUIPMENT/DEVICES	Flow Meter with Humidifier / Flow Meter / Oxygen Concentrator - Flow	Using Gas Flow Analyzer by Comparison Method	1 lpm to 15 lpm	5.2 %
7	MEDICAL DEVICES-DISCHARGE EQUIPMENT/DEVICES	Infusion Pump / Feeding Pump - Flow	Using Infusion Device Analyzer by Direct Method	5 ml/hr to 1000 ml/hr	4.28 %
8	MEDICAL DEVICES-DISCHARGE EQUIPMENT/DEVICES	Infusion Pump / Feeding Pump - Occlusion Pressure	Using Infusion Device Analyzer by Direct Method	0 to 1300 mmHg	2.07 %
9	MEDICAL DEVICES-DISCHARGE EQUIPMENT/DEVICES	Infusion Pump / Feeding Pump - Volume	Using Infusion Device Analyzer by Direct Method	1 ml to 400 ml	6.4 %
10	MEDICAL DEVICES-DISCHARGE EQUIPMENT/DEVICES	Nebulizer - Flow	Using Gas Flow Analyzer by Comparison Method	7 lpm to 10 lpm	5 %
11	MEDICAL DEVICES-DISCHARGE EQUIPMENT/DEVICES	Nebulizer - Pressure	Using Gas Flow Analyzer by Comparison Method	0 to 50 psi	4.97 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

PSG CENTRE FOR INDUSTRIAL RESEARCH AND DEVELOPMENT, A 115, PSG
COLLEGE OF TECHNOLOGY, AVINASHI ROAD, PEELAMEDU, COIMBATORE, TAMIL
NADU, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-4660

Page No

16 of 26

Validity

30/01/2026 to 29/01/2030

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
12	MEDICAL DEVICES-DISCHARGE EQUIPMENT/DEVICES	Pulse Oximeter - Heart Rate	Using SPO2 Functional Tester by Direct Method	30 bpm to 150 bpm	13.07 %
13	MEDICAL DEVICES-DISCHARGE EQUIPMENT/DEVICES	Pulse Oximeter - SPO2	Using SPO2 Functional Tester by Direct Method	80 % to 100 %	4.46 %
14	MEDICAL DEVICES-DISCHARGE EQUIPMENT/DEVICES	Suction Pump - Flow	Using Gas Flow Analyzer by Comparison Method	1 lpm to 15 lpm	2.41 %
15	MEDICAL DEVICES-DISCHARGE EQUIPMENT/DEVICES	Suction Pump - Vacuum	Using Gas Flow Analyzer by Comparison Method	0 to 600 mmHg	13.8 mmHg
16	MEDICAL DEVICES-DISCHARGE EQUIPMENT/DEVICES	Syringe Pump - Flow	Using Infusion Device Analyzer by Direct method	5 ml/hr to 1000 ml/hr	4.28 %
17	MEDICAL DEVICES-DISCHARGE EQUIPMENT/DEVICES	Syringe Pump - Occlusion Pressure	Using Infusion Device Analyzer by Direct Method	0 to 900 mmHg	4.29 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

PSG CENTRE FOR INDUSTRIAL RESEARCH AND DEVELOPMENT, A 115, PSG
COLLEGE OF TECHNOLOGY, AVINASHI ROAD, PEELAMEDU, COIMBATORE, TAMIL
NADU, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-4660

Page No

17 of 26

Validity

30/01/2026 to 29/01/2030

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
18	MEDICAL DEVICES-DISCHARGE EQUIPMENT/DEVICES	Syringe Pump - Volume	Using Infusion Device Analyzer by Direct Method	1 ml to 50 ml	6.4 %
19	MEDICAL DEVICES-DISCHARGE EQUIPMENT/DEVICES	Syringe Pump, Infusion Pump, Nebulizer, Pulse Oximeter, BP Apparatus, Suction Pump - Chassis/Enclosure Leakage Current	Using Electrical Safety Analyzer by Direct Method	10 μ A to 9000 μ A	3.6 μ A to 223 μ A
20	MEDICAL DEVICES-DISCHARGE EQUIPMENT/DEVICES	Syringe Pump, Infusion Pump, Nebulizer, Pulse Oximeter, BP Apparatus, Suction Pump - Earth Leakage Current	Using Electrical Safety Analyzer by Direct Method	10 μ A to 9000 μ A	3.6 μ A to 223 μ A
21	MEDICAL DEVICES-DISCHARGE EQUIPMENT/DEVICES	Syringe Pump, Infusion Pump, Nebulizer, Pulse Oximeter, BP Apparatus, Suction Pump - Insulation Resistance Test @ 500 V	Using Electrical Safety Analyzer by Direct Method	1 Mohm to 100 Mohm	8.98 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

PSG CENTRE FOR INDUSTRIAL RESEARCH AND DEVELOPMENT, A 115, PSG
COLLEGE OF TECHNOLOGY, AVINASHI ROAD, PEELAMEDU, COIMBATORE, TAMIL
NADU, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-4660

Page No

18 of 26

Validity

30/01/2026 to 29/01/2030

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
22	MEDICAL DEVICES-DISCHARGE EQUIPMENT/DEVICES	Syringe Pump, Infusion Pump, Nebulizer, Pulse Oximeter, BP Apparatus, Suction Pump - Patient Leakage Current	Using Electrical Safety Analyzer by Direct Method	10 μ A to 1000 μ A	3.6 μ A to 24.5 μ A
23	MEDICAL DEVICES-DISCHARGE EQUIPMENT/DEVICES	Syringe Pump, Infusion Pump, Nebulizer, Pulse Oximeter, BP Apparatus, Suction Pump - Protective Earth Resistance	Using Electrical Safety Analyzer by Direct Method	0.1 ohm to 1.9 ohm	0.022 ohm to 0.058 ohm
24	MEDICAL DEVICES-IMAGING/PLOTTERS	ECG Machine - Chassis / Enclosure Leakage Current	Using Electrical Safety Analyzer by Direct Method	10 μ A to 9000 μ A	3.6 μ A to 223 μ A
25	MEDICAL DEVICES-IMAGING/PLOTTERS	ECG Machine - Earth Leakage Current	Using Electrical Safety Analyzer by Direct Method	10 μ A to 10 mA	3.6 μ A to 223 μ A
26	MEDICAL DEVICES-IMAGING/PLOTTERS	ECG Machine - Insulation Resistance Test @ 500 V	Using Electrical Safety Analyzer by Direct Method	1 Mohm to 100 Mohm	3 %
27	MEDICAL DEVICES-IMAGING/PLOTTERS	ECG Machine - Patient Leakage Current	Using Electrical Safety Analyzer by Direct Method	10 μ A to 1000 μ A	3.6 μ A to 24.5 μ A



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

PSG CENTRE FOR INDUSTRIAL RESEARCH AND DEVELOPMENT, A 115, PSG
COLLEGE OF TECHNOLOGY, AVINASHI ROAD, PEELAMEDU, COIMBATORE, TAMIL
NADU, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-4660

Page No

19 of 26

Validity

30/01/2026 to 29/01/2030

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
28	MEDICAL DEVICES-IMAGING/PLOTTERS	ECG Machine - Protective Earth Resistance	Using Electrical Safety Analyzer by Direct Method	0.05 ohm to 1.9 ohm	0.022 ohm to 0.058 ohm
29	MEDICAL DEVICES-IMAGING/PLOTTERS	ECG Machine / Recorder - ECG Output Amplitude	Using Defibrillator Analyzer by Direct Method	0.5 mV to 5 mV	4.8 %
30	MEDICAL DEVICES-IMAGING/PLOTTERS	Heart Rate - ECG Machine/Recorder, Monitors	Using Defibrillator Analyzer by Direct Method	30 bpm to 300 bpm	2.27 %
31	MEDICAL DEVICES-MONITORING UNIT	Patient Monitor / Multipara Monitor - Chassis / Enclosure Leakage Current	Using Electrical Safety Analyzer by Direct Method	10 μ A to 9000 μ A	3.6 μ A to 223 μ A
32	MEDICAL DEVICES-MONITORING UNIT	Patient Monitor / Multipara Monitor - Earth Leakage Current	Using Electrical Safety Analyser by Direct Method	10 μ A to 9000 μ A	3.6 μ A to 223 μ A
33	MEDICAL DEVICES-MONITORING UNIT	Patient Monitor / Multipara Monitor - ECG	Using Vital Sign Simulator by Direct Method	0.05 mV to 5 mV	6.26 %
34	MEDICAL DEVICES-MONITORING UNIT	Patient Monitor / Multipara Monitor - Heart Rate	Using SPO2 Simulator and Vital Sign Simulator by Direct Method	30 BPM to 300 BPM	2.27 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

PSG CENTRE FOR INDUSTRIAL RESEARCH AND DEVELOPMENT, A 115, PSG
COLLEGE OF TECHNOLOGY, AVINASHI ROAD, PEELAMEDU, COIMBATORE, TAMIL
NADU, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-4660

Page No

20 of 26

Validity

30/01/2026 to 29/01/2030

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
35	MEDICAL DEVICES-MONITORING UNIT	Patient Monitor / Multipara Monitor - Invasive Blood Pressure Accuracy	Using SPO2 Simulator and Vital Sign Simulator by Direct Method	0 to 200 mmHg	5.33 %
36	MEDICAL DEVICES-MONITORING UNIT	Patient Monitor / Multipara Monitor - Non Invasive Blood Pressure	Using Vital Sign Simulator by Direct Method	50 to 400 mmHg	5.8 mmHg
37	MEDICAL DEVICES-MONITORING UNIT	Patient Monitor / Multipara Monitor - Non Invasive Blood Pressure (Leak Test)	Using Vital Sign Simulator by Direct Method	50 mmHg to 400 mmHg	1.3 mmHg
38	MEDICAL DEVICES-MONITORING UNIT	Patient Monitor / Multipara Monitor - Patient Leakage Current	Using Electrical Safety Analyzer by Direct Method	10 µA to 1000 µA	3.6 µA to 24.5 µA
39	MEDICAL DEVICES-MONITORING UNIT	Patient Monitor / Multipara Monitor - Protective Earth Resistance	Using Electrical Safety Analyzer by Direct Method	0.05 ohm to 1.9 ohm	0.022 ohm to 0.058 ohm
40	MEDICAL DEVICES-MONITORING UNIT	Patient Monitor / Multipara Monitor - Respiration Rate	Using Vital Sign Simulator by Direct Method	10 brpm to 100 brpm	6.6 %
41	MEDICAL DEVICES-MONITORING UNIT	Patient Monitor / Multipara Monitor - SPO2	Using SPO2 Simulator by Direct Method	70 % to 100 %	4.48 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

PSG CENTRE FOR INDUSTRIAL RESEARCH AND DEVELOPMENT, A 115, PSG
COLLEGE OF TECHNOLOGY, AVINASHI ROAD, PEELAMEDU, COIMBATORE, TAMIL
NADU, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-4660

Page No

21 of 26

Validity

30/01/2026 to 29/01/2030

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
42	MEDICAL DEVICES-MONITORING UNIT	Patient Monitor, Syringe / Infusion Pump, Defibrillator, Suction Pump, ECG Unit, Electro - Surgical Unit, Autoclave, Radiant Warmer, Patient Warmer, Infant Incubator - Insulation Resistance @ 500 V	Using Electrical Safety Analyzer by Direct Method	1 Mohm to 100 Mohm	8.98 %
43	MEDICAL DEVICES-PATIENT CONDITIONING / MAINTENANCE	Autoclave - Pressure	Using Digital Pressure Gauge by Comparison Method	0 to 2 bar	0.1 bar
44	MEDICAL DEVICES-PATIENT CONDITIONING / MAINTENANCE	Autoclave - Temperature	Using Digital Thermometer with Logger by Comparison Method	50 °C to 125 °C	1.3 °C
45	MEDICAL DEVICES-PATIENT CONDITIONING / MAINTENANCE	Autoclave - Time	Using Digital Timer by Comparison Method	1 s to 45 Min	0.1 s to 10 s



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

PSG CENTRE FOR INDUSTRIAL RESEARCH AND DEVELOPMENT, A 115, PSG
COLLEGE OF TECHNOLOGY, AVINASHI ROAD, PEELAMEDU, COIMBATORE, TAMIL
NADU, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-4660

Page No

22 of 26

Validity

30/01/2026 to 29/01/2030

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
46	MEDICAL DEVICES- PATIENT CONDITIONING / MAINTENANCE	Defibrillator - Charge Time	Using Defibrillator Analyzer by Direct Method	1.5 s to 10 s	0.17 s
47	MEDICAL DEVICES- PATIENT CONDITIONING / MAINTENANCE	Defibrillator - ECG Amplitude	Using Defibrillator Analyzer by Direct Method	0.5 mV to 5 mV	6.26 %
48	MEDICAL DEVICES- PATIENT CONDITIONING / MAINTENANCE	Defibrillator - Energy	Using Defibrillator Analyzer by Direct Method	20 J to 300 J	7.5 J
49	MEDICAL DEVICES- PATIENT CONDITIONING / MAINTENANCE	Defibrillator - Heart Rate	Using Defibrillator Analyzer by Direct Method	30 bpm to 300 bpm	1.6 % to 2.27 %
50	MEDICAL DEVICES- PATIENT CONDITIONING / MAINTENANCE	Defibrillator - NIBP	Using Vital Signs Simulator by Direct Method	50 mmHg to 200 mmHg	5.8 mmHg



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

PSG CENTRE FOR INDUSTRIAL RESEARCH AND DEVELOPMENT, A 115, PSG
COLLEGE OF TECHNOLOGY, AVINASHI ROAD, PEELAMEDU, COIMBATORE, TAMIL
NADU, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-4660

Page No

23 of 26

Validity

30/01/2026 to 29/01/2030

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
51	MEDICAL DEVICES- PATIENT CONDITIONING / MAINTENANCE	Defibrillator - SPO2	Using SPO2 Functional Tester by Direct Method	80 % to 100 %	4.46 %
52	MEDICAL DEVICES- PATIENT CONDITIONING / MAINTENANCE	Defibrillator, Infant Incubator, Electro Surgical Unit / Diathermy Machine / Cautery Machine, Autoclave, Patient Warmer, Radiant Warmer - Chassis / Enclosure Leakage Current	Using Electrical Safety Analyzer by Direct Method	10 μ A to 9000 μ A	3.6 μ A to 223 μ A
53	MEDICAL DEVICES- PATIENT CONDITIONING / MAINTENANCE	Defibrillator, Infant Incubator, Electro Surgical Unit / Diathermy Machine / Cautery Machine, Autoclave, Patient Warmer, Radiant Warmer - Earth Leakage Current	Using Electrical Safety Analyzer by Direct Method	10 μ A to 9000 μ A	3.6 μ A to 223 μ A



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

PSG CENTRE FOR INDUSTRIAL RESEARCH AND DEVELOPMENT, A 115, PSG
COLLEGE OF TECHNOLOGY, AVINASHI ROAD, PEELAMEDU, COIMBATORE, TAMIL
NADU, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-4660

Page No

24 of 26

Validity

30/01/2026 to 29/01/2030

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
54	MEDICAL DEVICES- PATIENT CONDITIONING / MAINTENANCE	Defibrillator, Infant Incubator, Electro Surgical Unit / Diathermy Machine / Cautery Machine, Autoclave, Patient Warmer, Radiant Warmer - Insulation Resistance Test @ 500 V	Using Electrical Safety Analyzer by Direct Method	1 Mohm to 100 Mohm	8.98 %
55	MEDICAL DEVICES- PATIENT CONDITIONING / MAINTENANCE	Defibrillator, Infant Incubator, Electro Surgical Unit / Diathermy Machine / Cautery Machine, Autoclave, Patient Warmer, Radiant Warmer - Patient Leakage Current	Using Electrical Safety Analyzer by Direct Method	10 μ A to 1000 μ A	3.6 μ A to 24.5 μ A
56	MEDICAL DEVICES- PATIENT CONDITIONING / MAINTENANCE	Defibrillator, Infant Incubator, Electro Surgical Unit / Diathermy Machine / Cautery Machine, Autoclave, Patient Warmer, Radiant Warmer - Protective Earth Resistance	Using Electrical Safety Analyser by Direct Method	0.1 ohm to 1.9 ohm	0.022 ohm to 0.058 ohm



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

PSG CENTRE FOR INDUSTRIAL RESEARCH AND DEVELOPMENT, A 115, PSG
COLLEGE OF TECHNOLOGY, AVINASHI ROAD, PEELAMEDU, COIMBATORE, TAMIL
NADU, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-4660

Page No

25 of 26

Validity

30/01/2026 to 29/01/2030

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
57	MEDICAL DEVICES- PATIENT CONDITIONING / MAINTENANCE	Electro Surgical Unit / Diathermy Machine / Cautery Machine - HF Leakage	Using Electro Surgery Analyzer by Direct Method	11 mA to 150 mA	1.62 mA to 5.64 mA
58	MEDICAL DEVICES- PATIENT CONDITIONING / MAINTENANCE	Electro Surgical Unit / Diathermy Machine / Cautery Machine - Power	Using Electro Surgery Analyzer by Direct Method	10 W to 300 W	8.48 %
59	MEDICAL DEVICES- PATIENT CONDITIONING / MAINTENANCE	Electro Surgical Unit / Diathermy Machine / Cautery Machine - REM	Using Electro Surgery Analyzer by Direct Method	10 ohm to 475 ohm	3.08 %
60	MEDICAL DEVICES- PATIENT CONDITIONING / MAINTENANCE	Pacemaker / Defibrillator - Pacer Output	Using Defibrillator analyzer by Direct Method	2 mA to 20 mA	3.8 %
61	MEDICAL DEVICES- PATIENT CONDITIONING / MAINTENANCE	Pacemaker / Defibrillator - Pacer Rate	Using Defibrillator analyzer by Direct Method	30 ppm to 180 ppm	1.92 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

PSG CENTRE FOR INDUSTRIAL RESEARCH AND DEVELOPMENT, A 115, PSG
COLLEGE OF TECHNOLOGY, AVINASHI ROAD, PEELAMEDU, COIMBATORE, TAMIL
NADU, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-4660

Page No

26 of 26

Validity

30/01/2026 to 29/01/2030

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)($\pm$)
62	MEDICAL DEVICES- PATIENT CONDITIONING / MAINTENANCE	Patient Warmer - Temperature	Using Digital Thermometer with Logger by Comparison Method	30 °C to 45 °C	1 °C
63	MEDICAL DEVICES- PATIENT CONDITIONING / MAINTENANCE	Radiant Warmer - Temperature	Using Digital Thermometer with Logger by Comparison Method	30 °C to 40 °C	1.2 °C

* CMCs represent expanded uncertainties expressed at approximately the 95% level of confidence, using a coverage factor of $k = 2$.